

The Tracoe Twist Plus P-tube premounted on its insertion system for a minimally traumatic tube placement.



For the Tracoe Twist Plus P-tubes, the Experc dilator and guiding catheter are removed after the tracheotomy so that only the guide wire remains in the patient.



Check the cuff for leaks, then deflate and lubricate the tube end, silicone sleeve and cuff material with lubricating jelly before placement.



Introduce the insertion system with the Twist Plus tube with pre-mounted guiding catheter onto the guide wire.



Do not disassemble the pre-assembled tracheostomy tube consisting of insertion system with guiding catheter, tracheostomy tube and inner cannula.



Bring the tip of the insertion system to skin level using the safety stopper of the guiding catheter.



Visualize the mark at the end of the guide wire.



Pinch the guide wire with guiding catheter at the end of the insertion system into the notch provided.



Start the insertion vertically and place the tube in a controlled movement until the neck flange reaches skin level.



Hold the neck flange with two fingers and remove the inserter, guiding catheter and Seldinger guide wire.



Tip: Do not pull on the tip of the guiding catheter or silicone sleeve solely. Ensure to take hold of the complete insertion system (inserter and guiding catheter).





While pulling the insertion system out, the silicone sleeve automatically flips over and can therefore be removed.



Inflate the cuff of the tracheostomy tube, disconnect the ventilation tube from the ET tube and attach it to the 15 mm connector of the already situated, non-fenestrated inner cannula.

Ensure that the ventilator circuit has been disconnected from the ET tube before inflating the tracheostomy tube cuff.



The position of the tracheostomy tube is confirmed via the ventilatory circuit using a fiber optic camera. In addition, another fiberoptic assessment is carried out via the ET tube.

If necessary, correct the position of the tracheostomy tube during the check-ups.



The cuff of the ET tube is then deflated, withdrawn and removed under control of the anatomical structures.

The ventilation parameters must then be adjusted.



Place a dressing under the neck flange and fix the tube in place with the neck strap.

Tip: Ongoing evaluation of tube position is recommended throughout patient journey.



Now the cuff pressure should be checked, for example with a handheld manometer. It should be between 20 and 30 cmH₂O.

16

Alternatively, our Tracoe Smart Cuff Manager can be connected for continuous monitoring and regulation of the recommended cuff pressure.

© Coloplast A/S. All rights reserved. Atos and the Atos Medical logo are trademarks of Coloplast A/S. Please also note the applicable instructions for use. These instructions do not replace the IF

Head office: Atos Medical AB Hyllie Boulevard 17 SE-215 32 Malmo, Sweden Tel: +46 (0) 415 198 00 Email: info@atosmedical.com Manufacturer: Tracoe Medical GmbH Reichelsheimer Straße 1/3 55268 Nieder-Olm, Germany



MC2998-TcEN 202304